

**EUROPEAN AND MEDITERRANEAN PLANT PROTECTION ORGANIZATION**  
**ORGANISATION EUROPEENNE ET MEDITERRANEENNE POUR LA PROTECTION DES PLANTES**  
**Summary sheet of validation data for a diagnostic test**

The EPPO Standard PM 7/98 *Specific requirements for laboratories preparing accreditation for a plant pest diagnostic activity* describes how validation should be conducted. It also includes definitions of performance criteria.

<b>Laboratory contact details</b>	EUPHRESKO-GRAFDEPI Via Carlo Giuseppe Bertero, 22, 00156 ROMA, Italy
<b>Short description of the test</b>	Simplex real time PCR for the detection of FD and BN phytoplasmas with an internal control for grapevine - Angelini et al., 2007.
<b>Date, reference of the validation report</b>	2014-07-31 - Project EUPHRESKO GRAFDEPI Final Report - 1) <a href="http://www.euphresco.net/media/project_reports/grafdepi_final_report.pdf">http://www.euphresco.net/media/project_reports/grafdepi_final_report.pdf</a> 2) The Euphresco Grafdepi Group, 2015. European interlaboratory comparison of detection methods for "flavescence dorée" phytoplasma: preliminary results. Phytopathogenic Mollicutes doi: 10.5958/2249-4677.2015.00015.8 Vol. 5 (1-Supplement), January 2015, S35-S37
<b>Validation process according to EPPO Standard PM7/98?</b>	yes
<b>Is the lab accredited for this test?</b>	no
<b>Was the validated data generated in the framework of a project?</b>	Euphresco
<b>If yes, please specify</b>	GRAFDEPI
<b>Description of the test</b>	
<b>Organism(s)</b>	Grapevine flavescence dorée phytoplasma (PHYP64)
<b>Detection / identification</b>	detection
<b>Method(s)</b>	Molecular real time PCR
<b>Method: Molecular real time PCR</b>	
<b>Reference of the test description</b>	
<b>As or adapted from an EPPO diagnostic protocol</b>	no
<b>As or adapted from an IPPC diagnostic protocol</b>	no
<b>Reference of the test</b>	ANGELINI E., BIANCHI G.L., FILIPPIN L., MORASSUTTI C., BORGO M. A new TaqMan method for the identification of phytoplasmas associated with grapevine yellows by real-time PCR assay. Journal of Microbiological Methods, 68 (2007), 613-622

<b>Other information</b>	
<b>Reaction type</b>	Simplex - Probe
<b>Other details on the test</b>	A TaqMan real time PCR methodology performed with specific primers and probes for the detection of FD and BN phytoplasmas and a grapevine internal control (Grapevine chloroplast chaperonin 21 gene). Detection of the single targets must be performed in separate reactions.
<b>Are the performance characteristics included in the EPPO diagnostic protocol?</b>	<b>no</b>
<b>Performance Criteria :</b>	
<b>Organism 1.:</b>	<b>Grapevine flavescence dorée phytoplasma(PHYP64)</b>
<b>Analytical sensitivity</b>	
<b>What is smallest amount of target that can be detected reliably?</b>	The analytical sensitivity was calculated in five laboratories analyzing three samples at five dilution levels (1/10; 1/100; 1/300; 1/900; 1/2700) in five repetitions. Samples (DNA extracts) came from IPEP (Serbia), ACW (Switzerland) and ANSES (France). The homogenising and preparation were performed by ANSES-LSV (France). Two values are provided: The last dilution level with 100% positive results: less than 1/10 (for all samples) The last dilution level with, at least, one positive result for each sample: 1/2700 (for all samples)
<b>Diagnostic sensitivity</b>	
<b>Proportion of infected/infested samples tested positive compared to results from the standard test, see appendix 2 of PM 7/98</b>	Seven laboratories performed this protocol within GRAFDEPI ringtest with a total of 168 results. Determined in 11 samples positive for Flavescence dorée phytoplasma. They were DNA extracts of Vitis sp. tested positive by PCR for FD pure or mixed with different quantities of healthy grapevine or mixed with DNA extracts positive for the 16SrXII group phytoplasmas. Within the ringtest 7 diagnostic methods were compared. Diagnostic sensitivity: 86.67% False negative: (14/168) 8.3%
<b>Standard test(s)</b>	Other protocols included in the ringtest: - Simultaneous detection of FD and BN phytoplasmas by multiplex nested-PCR (Dairè et al., 1997; Angelini et al., 2001; Clair et al., 2003) - Detection of Flavescence dorée phytoplasma by universal direct PCR and nested 16SrV-group specific PCR - Detection and identification of Flavescence dorée phytoplasma by direct and nested PCR followed by RFLP with Taq I (Martini et al., 1999) - Simplex real time PCR for the detection of FD and BN phytoplasmas with an internal control - (Hren et al., 2007) - Triplex real-time PCR for simultaneous FD and BN phytoplasmas detection with an internal control for grapevine. (Pelletier et al., 2009) - Triplex real time PCR for simultaneous FD and BN phytoplasmas detection with an internal

	control - (under patent IPADLAB)
<b>Diagnostic Specificity</b>	
<b>Proportion of uninfected/uninfested samples (true negatives) testing negative compared to results from a standard test</b>	Seven laboratories performed this protocol within GRAFDEPI ringtest with a total of 168 results. 13 non target samples: 4 healthy grapevines and 9 were other phytoplasmas of 16SrV group and phytoplasmas from other groups. Diagnostic specificity: 66.10%
<b>Specify the test(s)</b>	Other protocols included in the ringtest: - Simultaneous detection of FD and BN phytoplasmas by multiplex nested-PCR (Dairè et al., 1997; Angelini et al., 2001; Clair et al., 2003) - Detection of Flavescence dorée phytoplasma by universal direct PCR and nested 16SrV-group specific PCR - Detection and identification of Flavescence dorée phytoplasma by direct and nested PCR followed by RFLP with Taq I (Martini et al., 1999) - Simplex real time PCR for the detection of FD and BN phytoplasmas with an internal control - (Hren et al., 2007) - Triplex real-time PCR for simultaneous FD and BN phytoplasmas detection with an internal control for grapevine. (Pelletier et al., 2009) - Triplex real time PCR for simultaneous FD and BN phytoplasmas detection with an internal control - (under patent IPADLAB)
<b>Reproducibility</b>	
<b>Provide the calculated % of agreement for a given level of the pest (see PM 7/98)</b>	The reproducibility was calculated in five laboratories analyzing three samples at five dilution levels (1/10; 1/100; 1/300; 1/900; 1/2700) in five repetitions. Samples (DNA extracts) came from IPEP (Serbia), ACW (Switzerland) and ANSES (France). The homogenising and preparation were performed by ANSES-LSV (France). Reproducibility: 75.59%
<b>Repeatability</b>	
<b>Provide the calculated % of agreement for a given level of the pest (see PM 7/98)</b>	The repeatability was calculated in five laboratories analysing three samples at five dilution levels (1/10; 1/100; 1/300; 1/900; 1/2700) in five repetitions. Samples (DNA extracts) came from IPEP (Serbia), ACW (Switzerland) and ANSES (France). The homogenising and preparation were performed by ANSES-LSV (France) Repeatability: 88.05%
<b>Test performance study</b>	
<b>Test performance study?</b>	yes
<b>Brief details of the test performance study and its output. It available, link to published article/report</b>	Interlaboratory comparison among 15 laboratories within the EUPHRESKO Project GRAFDEPI (CRA-PAV, Italy; AGES, Austria; CRA-W, Belgium, PPRS, Turkey; INIAV, Portugal; ACW, Switzerland; ILVO, Belgium; DIPSA, Bologna Italy; DISAA, Milan Italy; IPEP, Serbia; NIB, Slovenia; IRTA, Spain; ANSES, France; Cra-VIT, Italy)
<b>Other information</b>	

<b>Any other information considered useful</b>	The ringtest was carried out by 15 laboratories and it is not possible to state if any of them is accredited for this test.
The following complementary files are available online:	<ul style="list-style-type: none"> <li>• <a href="#">EUPHRESKO-GRAFDEPI Samples for determination of performance criteria</a></li> </ul>

*Creation date: 2015-02-10 00:00:00 - Last update: 2021-05-17 17:10:08*