

**EUROPEAN AND MEDITERRANEAN PLANT PROTECTION ORGANIZATION**  
**ORGANISATION EUROPEENNE ET MEDITERRANEENNE POUR LA PROTECTION DES PLANTES**  
**Summary sheet of validation data for a diagnostic test**

The EPPO Standard PM 7/98 *Specific requirements for laboratories preparing accreditation for a plant pest diagnostic activity* describes how validation should be conducted. It also includes definitions of performance criteria.

<b>Laboratory contact details</b>	ILVO Institute for Agricultural and Fisheries Research Burg. Van Gansberghelaan 96, 9820 Merelbeke - Melle, Belgium
<b>Short description of the test</b>	Identification of Meloidogyne graminicola by molecular conventional PCR Bellafigliore et al. 2015 in juveniles
<b>Date, reference of the validation report</b>	2023-07-31 - TEST PERFORMANCE STUDY REPORT 22MG
<b>Link to other validation data</b>	- TEST PERFORMANCE STUDY REPORT 22MG Identification of Meloidogyne graminicola by molecular conventional PCR Htay et al 2016 in juveniles - TEST PERFORMANCE STUDY REPORT 22MG Identification of Meloidogyne graminicola by molecular conventional PCR Mattos et al 2019 (oryzae primers) in juveniles - Validation report for the molecular identification of Meloidogyne graminicola Identification of Meloidogyne graminicola by molecular conventional PCR Htay et al., 2016 in juveniles - Validation report for the molecular identification of Meloidogyne graminicola Identification of Meloidogyne graminicola by molecular conventional PCR Bellafigliore et al. 2015 in juveniles
<b>Validation process according to EPPO Standard PM7/98?</b>	yes
<b>Is the lab accredited for this test?</b>	no
<b>Was the validated data generated in the framework of a project?</b>	EURL
<b>If yes, please specify</b>	EU-funded project EURLs-EURCs 2021-2022 (grant SI2.870859)
<b>Description of the test</b>	
<b>Organism(s)</b>	Meloidogyne graminicola (MELGGC)
<b>Detection / identification</b>	identification
<b>Method(s)</b>	Molecular Extraction DNA RNA Molecular Conventional PCR Molecular Conventional PCR (2)

<b>Method: Molecular Extraction DNA RNA</b>	
<b>Reference of the test description</b>	
<b>Kit</b>	
Is a kit used	yes
Manufacturer name	
Specify the kit used	
Kit used following the manufacturer's instructions?	
<b>Other information</b>	
Other details on the test	Check TPS report
<b>Method: Molecular Conventional PCR</b>	
<b>Reference of the test description</b>	
As or adapted from an EPPO diagnostic protocol	no
New test being considered for inclusion in the next version of the EPPO diagnostic protocol?	yes
As or adapted from an IPPC diagnostic protocol	no
Reference of the test	Mattos et al., 2019
<b>Other information</b>	
Reaction type	Simplex
Other details on the test	M. oryzae primers
<b>Method: Molecular Conventional PCR (2)</b>	
<b>Reference of the test description</b>	
As or adapted from an EPPO diagnostic protocol	no
New test being considered for inclusion in the next version of the EPPO diagnostic protocol?	yes
As or adapted from an IPPC diagnostic protocol	no
Reference of the test	Bellafigliore et al. 2015
<b>Kit</b>	
Is a kit used	no
<b>Other information</b>	
Reaction type	Simplex
Other details on the test	SCAR primers
Are the performance characteristics included in the EPPO diagnostic protocol?	no
<b>Performance Criteria :</b>	
Organism 1.:	Meloidogyne graminicola(MELGGC)

<b>Analytical sensitivity</b>	
<b>What is smallest amount of target that can be detected reliably?</b>	Analytical sensitivity for 1 nematode: amplification in 6 out of 18 replicates: 33% Analytical sensitivity for 2 nematodes: amplification in 8 out of 18 replicates: 44% Analytical sensitivity for 5 nematodes: amplification in 10 out of 18 replicates: 56% Analytical sensitivity for 10 nematodes: amplification in 16 out of 18 all replicates: 89%
<b>Diagnostic sensitivity</b>	
<b>Proportion of infected/infested samples tested positive compared to results from the standard test, see appendix 2 of PM 7/98</b>	Results from 7 laboratories when the test was used in combination with Mattos (2019) M.oryzae primers: DSE=98%
<b>Analytical specificity - inclusivity</b>	
<b>Number of strains/populations of target organisms tested</b>	Population from Italy and the Philippines amplified
<b>Specificity value</b>	
<b>Analytical specificity - exclusivity</b>	
<b>Number of non-target organisms tested</b>	TPS: M. incognita, M. naasi, M. oryzae
<b>Specificity value</b>	
<b>Cross reacts with</b>	Meloidogyne oryzae
<b>Diagnostic Specificity</b>	
<b>Proportion of uninfected/uninfested samples (true negatives) testing negative compared to results from a standard test</b>	Results from 7 laboratories when the test was used in combination with Mattos (2019) M.oryzae primers: DSP=100%
<b>Test performance study</b>	
<b>Test performance study?</b>	yes
<b>Brief details of the test performance study and its output.It available, link to published article/report</b>	TPS studies involving 9 laboratories, 6 target samples (2 populations, 3 samples per populations), 9 non target samples (3 samples for each 3 species M nassi, M. oryzae and M incognita).
<b>Other information</b>	
<b>Any other information considered useful</b>	TPS report available on the EURL website: <a href="https://sites.anses.fr/en/system/files/TestPerformanceStudy_Report_Meloidogyne_graminicola.pdf">https://sites.anses.fr/en/system/files/TestPerformanceStudy_Report_Meloidogyne_graminicola.pdf</a>

Creation date: 2023-11-14 09:30:56 - Last update: 2023-11-14 14:14:00